



NÁRODNÍ AKREDITAČNÍ ORGÁN

EA MLA Signatory  
Český institut pro akreditaci, o.p.s.  
Olšanská 54/3, 130 00 Praha 3

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products, as amended

## CERTIFICATE OF ACCREDITATION

No. 514/2018

**František Knížek**  
with registered office A. Dvořáka 609, 533 41 Lázně Bohdaneč, Company Registration  
No. 46494111

to the Calibration Laboratory No. 2290  
František Knížek – KALEX, Calibration Centre

Scope of accreditation:

Calibration of meters in the field of length, plane angle, mass, force and torque, pressure and temperature to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of Accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

ČSN EN ISO/IEC 17025:2018

In its activities performed within the scope and for the period of validity of this Certificate, the Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited Conformity Assessment Body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 427/2017 of 18. 7. 2017, or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **1. 10. 2023**

Prague: 1. 10. 2018



*Jiří Růžička*  
**Jiří Růžička**  
Director  
Czech Accreditation Institute  
Public Service Company

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

**Calibration laboratory locations:**

1. **Workplace Lázně Bohdaneč** A. Dvořáka 719, 533 41 Lázně Bohdaneč
2. **Workplace Vlčí Habřina** Vlčí Habřina 122, 533 41 Lázně Bohdaneč

**CMC for the field of measured quantity: Length**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workpl ace
		min.	unit	max.	unit					
1	Steel parallels	0.5 mm	up to	100 mm			0.2+2L μm	Comparison with the standard	KPA-1.01	1
		125 mm	up to	500 mm			0.3+2.2L μm			1, 2
		500 mm	up to	1,000 mm			0.3+2.2L μm			2
2*	Slide gauges, depth gauges, height gauges	0 mm	up to	3,000 mm			11+8.7L μm	Comparison with the standard	KPA-1.02	1
3	Micrometers for external measurement							Comparison with the standard	KPA-1.03	1
	Micrometer calliper gauges	0 mm	up to	500 mm			1+3L μm			
	Pasameters	0 mm	up to	500 mm			1+3L μm			
	Micropasameters	0 mm	up to	500 mm			1+3L μm			
4	Micrometers for internal measurement							Comparison with the standard	KPA-1.04	1
	Inside micrometer gauges	14 mm	up to	500 mm			1+3L μm			
	Micrometer depth gauges	14 mm	up to	500 mm			1.1+2L μm			
	Inside micrometers	14 mm	up to	500 mm			1.1+2L μm			
	Micrometric heads	0 mm	up to	500 mm			1+3L μm			
5	Deviation meters							Direct measurement	KPA-1.05	1
	Dial gauges	0 mm	up to	100 mm			0.88 μm			
	Pupitasts	0 mm	up to	100 mm			1.2 μm			
	Somcators	0 mm	up to	100 mm			1.2 μm			
	Internal gauges	0 mm	up to	100 mm			1.2 μm			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workpl ace
		min.	unit	max.	unit					
6	Limit gauges for external measurement Micrometer calliper gauges Plain rings Threaded rings	1 mm	up to	500 mm			1+3L $\mu\text{m}$ 1.2+3.6L $\mu\text{m}$ 2.1+7.4L $\mu\text{m}$	Direct measurement	KPA-1.06	1
7	Limit gauges for internal measurement cylinder, flat Thread gauges Feeler gauges Measuring wires Gauges for radius Gauges for threads Gauges for paint thickness	0.05 mm	up to	500 mm			0.75+5.3L $\mu\text{m}$ 2.8+2.8L $\mu\text{m}$ 3.6 $\mu\text{m}$ 0.54 $\mu\text{m}$ 4.0 $\mu\text{m}$ 4.0 $\mu\text{m}$ 1.4 $\mu\text{m}$	Direct measurement	KPA-1.07	1
8*	Rules  Steel rules Measuring magnifier Steel tape measures Tapes	0 mm	up to	10,000 mm			4.7+4.6L $\mu\text{m}$ 4.7+4.6L $\mu\text{m}$ 140+3.8L $\mu\text{m}$ 0.3+0.06L $\mu\text{m}$	Comparison with the standard	KPA-1.08	1, 2
9*	Two-coordinate measuring machines, Measuring microscopes, Profile projectors	0 mm	up to	1,000 mm			3.2 $\mu\text{m}$	Comparison with the standard	KPA-1.09	1
10	Atypical length gauges	0 mm	up to	250 mm			4.0 $\mu\text{m}$	Measurement on a coordinate measuring machine	KPA-1.10	1
11*	Surface plates, blocks, plates	0 m	up to	5 m			3.8 $\mu\text{m}$	Direct measurement	KPA-1.13	1
12*	Length measuring instrument	0 mm	up to	1,000 mm			0.25+2L $\mu\text{m}$	Comparison with the standard	KPA-1.14	1

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

- <sup>1)</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- <sup>2)</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- <sup>3)</sup> If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).

L..Nominal length in metres



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

**CMC for the field of measured quantity: Plane angle**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workpl ace
		min.	unit	max	unit					
1	Rigid angle gauges							Direct measurement	KPA-1.11	1
	Check squares	0 °		up to	90 °		32 µm/m			
	Taper gauges	0 °		up to	90 °		7 "			
	Accuracy of levels	0 °		up to	90 °		5 µm/m			
	Centre square	0 °		up to	90 °		32 µm/m			
	Gauges for threads	0 °		up to	90 °		7 "			
2	Angle gauges							Comparison with the standard	KPA-1.12	1
	mechanical, digital, optical, with a dial indicator	0 °		up to	90 °		1.8 '			
	arc-shape	0 °		up to	90 °		0.7 °			

<sup>1)</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2)</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3)</sup> If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

**CMC for the field of measured quantity: Mass**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workpl ace
		min.	unit	max.	unit					
1*	Scales with non-automatic function, mechanical, digital	0 kg	up to	2 kg			$5 \cdot 10^{-6}$	Comparison with the standard	KPA-2.01	1
		2 kg	up to	3 kg			$1.6 \cdot 10^{-5}$	class E2		
		3 kg	up to	45 kg			$1.6 \cdot 10^{-5}$	class F1		
		45 kg	up to	6,000 kg			$5 \cdot 10^{-5}$	class F2		
								class M1		
2	Weights and other objects	6,000 kg	up to	30,000 kg			$1.6 \cdot 10^{-4}$	class M1 with substitute load	KPA-2.01	2
		1 g	up to	500 g			8.2 mg	Comparison with the standard		
		0.5 kg	up to	1 kg			8.6 mg			
		1 kg	up to	2 kg			10 mg			
		2 kg	up to	5 kg			16 mg			
		5 kg	up to	20 kg			59 mg			

<sup>1)</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2)</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3)</sup> If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

**CMC for the field of measured quantity: Force, mechanical tests**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workpl ace
		min.	unit	max.	unit					
1*	Torque wrenches, screwdrivers, Moment of force meters	0.25 Nm	up to	0.5 Nm			0.01	Comparison with the standard	KPA-5.01	1, 2
		0.5 Nm	up to	200 Nm			0.005			
		200 Nm	up to	2,000 Nm			0.005			
2*	Dynamometers, force measuring devices	0 N	up to	500 N		Tension, Pressure	0.001	Comparison with the standard	KPA-5.02	1
		500 N	up to	10,000 N			0.003			

<sup>1)</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2)</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3)</sup> If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).



**The Appendix is an integral part of  
Certificate of Accreditation No. 514/2018 of 01/10/2018**

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

**CMC for the field of measured quantity: Pressure, mechanical stress**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workpl ace
		min.	unit	max	unit					
1*	Deformation manometers, Digital manometers, Pressure measuring chains, Pressure transducers with electrical output	-95 kPa	up to	350 kPa		Underpressure/o verpressure      Gases	0.26 kPa	Comparison with the standard	KPA-4.01, KPA-4.02	1
		350 kPa	up to	1,000 kPa			0.58 kPa			
		1 MPa	up to	3.5 MPa			2.1 kPa			
		3.5 MPa	up to	6 MPa			6.9 kPa			
		0 MPa	up to	20 MPa		Liquids	35 kPa			
		20 MPa	up to	50 MPa			87 kPa			

<sup>1)</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2)</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3)</sup> If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

**CMC for the field of measured quantity: Temperature**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workpl ace
		min.	unit	max	unit					
1	Glass thermometers	- 40 °C	up to	200 °C	0 °C		0.07 °C 0.05 °C	Comparison with the standard	KPA-3.01	2
2*	Direct indicating thermometers, temperature controller	-40 °C 200 °C 400 °C 650 °C 900 °C	up to up to up to up to up to	200 °C 400 °C 650 °C 900 °C 1200 °C			0.08 °C 0.44 °C 1.5 °C 1.8 °C 2.4 °C	Comparison with the standard	KPA-3.02	1
3*	Infrared thermometers	50 °C	up to	500 °C			3.2 °C	Comparison with the standard	KPA-3.03	1
4*	Contact thermometers	0 °C 50 °C 100 °C 200 °C 400 °C	up to up to up to up to up to	50 °C 100 °C 200 °C 400 °C 600 °C			1.7 °C 1.9 °C 2.3 °C 2.6 °C 3.5 °C	Comparison with the standard	KPA-3.04	1
5*	Thermoelectric sensors and measuring chains Thermocouple sensors Measuring chain without a sensor	-40 °C 200 °C 400 °C 650 °C 900 °C -100 °C	up to up to up to up to up to up to	200 °C 400 °C 650 °C 900 °C 1,100 °C 1,100 °C		K, J, N	0.4 °C 0.6 °C 1.6 °C 2.3 °C 2.6 °C 0.3 °C	Comparison with the standard	KPA-3.05	1
6*	Resistance sensors and measuring chains Resistance sensors Measuring chain without a sensor	-40 °C 200 °C -100 °C	up to up to up to	200 °C 400 °C 400 °C			0.2 °C 0.5 °C 0.2 °C	Comparison with the standard	KPA-3.06	1

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**František Knížek**  
František Knížek – KALEX, Calibration Centre  
Antonína Dvořáka 719, 533 41 Lázně Bohdaneč

- <sup>1)</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- <sup>2)</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- <sup>3)</sup> If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).

